

Cambridge Low Contrast Gratings

A.J. Wilkins

Medical Research Council
Applied Psychology Unit, Cambridge

J.G. Robson

Physiological Laboratory
University of Cambridge

Clement Clarke International Ltd.

Edinburgh Way
Harlow

Essex CM20 2TT England

Telephone: (+44) 01279 414969 Telefax: (+44) 01279 635232

Catalogue no: 7002200

Issue no: 5/01

 **CLEMENT CLARKE
INTERNATIONAL**

Precision by Tradition

Conversion Table

Total
Score Contrast
Sensitivity

4.....	10
5.....	13
6.....	16
7.....	20
8.....	24
9.....	28
10.....	33
11.....	37
12.....	43
13.....	49
14.....	55
15.....	62
16.....	70
17.....	78
18.....	88
19.....	99
20.....	110
21.....	120
22.....	130
23.....	140
24.....	150
25.....	170
26.....	180
27.....	190
28.....	210
29.....	230
30.....	250
31.....	270
32.....	290
33.....	310
34.....	340
35.....	370
36.....	400
37.....	440
38.....	480
39.....	520
40.....	560

the lines 0.4 cpcm
1 cycle = 2.54 cm

Distance	cpd
57.2 cm	0.4 cpd
114 cm	0.8 cpd
171 cm	1.2 cpd
229 cm	1.6 cpd
286 cm	2.0 cpd
343 cm	2.4 cpd
400 cm	2.8 cpd
500 cm	3.3 cpd